



BG Analytics®

System Verification Protocol

G_1866 Rev3 2022-08-06

REF BGA007



IVD

This product is for In Vitro Diagnostic Use and Professional Use only
Visit www.acciusa.com for instructions for use in your language



ASSOCIATES OF
CAPE COD
INCORPORATED

124 Bernard E. Saint Jean Drive • E. Falmouth, MA 02536 USA

Telephone: (508) 540-3444
Toll-Free: (888) 395-2221
Fax: (508) 540-8680
Technical Support: (800) 848-3248
Customer Service: (800) 525-8378

This document has been prepared for use by the customers and authorized personnel of Associates of Cape Cod, Inc. The information contained in this manual is proprietary. The manual may not be copied, reproduced, translated or transmitted in any form without the express written permission of Associates of Cape Cod, Inc.

No commercial warranties of any kind, expressed or implied, are made.

G_1866 Rev3

Microsoft®, Microsoft® .NET, Windows® 10 are registered trademarks of Microsoft Corporation in the United States and/or other countries.

Windows® and the Windows logo are trademarks of the Microsoft group of companies.

BG Analytics® and Fungitell STAT® are registered trademarks of Associates of Cape Cod, Inc.

Lab Kinetics Incubating 8-well Tube Reader is an instrument manufactured by Lab Kinetics LLC.

Contents

1	Overview	5
1.1	Purpose	5
2	System Verification Plan	6
2.1	Scope	6
2.2	Description of the Components to Be Tested	6
2.3	Required Specifications	8
2.4	Laboratory's System Verification Plan	8
2.5	Responsibilities	9
2.6	List of Supplies Required to Execute this System Verification Protocol	12
2.7	Procedure	12
2.8	Acceptance Criteria	13
2.9	Location of the Completed System Verification Protocol	13
2.10	Review and Approval	14
3	Installation Qualification of Lab Kinetics Incubating 8-well Tube Reader	15
3.1	Calibration Documentation Test Case	15
3.2	Setup of Lab Kinetics Incubating 8-well Tube Reader Test Case	16
3.3	The Installation Lab Kinetics Incubating 8-well Tube Reader Test Case	17
3.4	Evaluation of Lab Kinetics Incubating 8-well Tube Reader Test Case	18
3.5	Verification of the Performance of Lab Kinetics Incubating 8-well Tube Reader Test Case	19
3.6	Review and Approval	21
4	Installation Qualification of BG Analytics® Software	22
4.1	Installation of BG Analytics® Software Test Case	22
4.2	Installation of Barcode Scanner Test Case	24
4.3	Review and Approval	25
5	Operational Qualification of Lab Kinetics Incubating 8-well Tube Reader and BG Analytics® Software	26
5.1	Verification of Data Transmission Test Case	26
5.2	Verification of Collecting, Saving, Analysis and Test Result Delivery Test Case	28
5.3	Verification of BG Analytics® Reporting Test Results Test Case	31
5.4	Verification of Data Storage and Searching Capabilities Test Case	34
5.5	Verification of Database Backup Capability Test Case	35
5.6	Review and Approval	36

6 Final Verification Report 37

6.1 Final Verification Report 37

6.2 Review and Approval 38

7 Attachments..... 39

7.1 Training Records 39

7.2 Objective Evidence..... 40

7.3 Additional Testing 41

7.4 Discrepancy Report..... 42

7.5 Problem Resolution Report..... 43

7.6 Maintenance 44

1 Overview

1.1 Purpose

This System Verification Protocol is designed to confirm that the system (where the system consists of Lab Kinetics Incubating 8-well Tube Reader and BG Analytics® software installed on a host computer) performs its required functions accurately and reliably. Specifically, the individual Test Cases (also referred to as TC number) outlined in this System Verification Protocol are designed to demonstrate, document, evaluate and confirm that the system performs as intended.

The following product abbreviations are used throughout this Protocol:

- **The PKF08 instrument (or PKF08)** for Lab Kinetics Incubating 8-well Tube Reader
- **BGA or BG Analytics®** for BG Analytics® Software
- **Fungitell STAT®** for Fungitell STAT® (1,3)-B-D-Glucan Detection Assay

Translated versions of this System Verification Protocol are available for download at: www.acciusa.com.

2 System Verification Plan

2.1 Scope

The scope of the System Verification Protocol for PKF08 and BGA defines the process by which the PKF08 instrument and BG Analytics® software will be verified for their intended purpose. The Required Specifications specify the purpose and functions required of the instrument and software as defined in the user needs. This System Verification Protocol specifies that each Required Specification be tested per pre-defined Test Cases included in the Installation Qualification and Operational Qualification Sections. Each executed Test Case includes a formal record of Expected and Observed Results. The Final Verification Report provides an overview of the status of the executed Test Cases and formally documents whether the system conforms to the Required Specifications.

This System Verification Protocol is divided into the following sections:

- **Section 3 Installation Qualification of Lab Kinetics Incubating 8-well Tube Reader** confirms that the PKF08 instrument is installed according to the manufacturer's specifications and the functional tests have been performed and documented with expected results.
- **Section 4 Installation Qualification of BG Analytics® Software** confirms that the software is installed according to the manufacturer's specifications and the functional tests have been performed and documented with expected results.
- **Section 5 Operational Qualification of Lab Kinetics Incubating 8-well Tube Reader and BG Analytics® Software** confirms that the system is operating within established limits and tolerances.
- **Section 6 Final Verification Report** provides an overview of applicable Test Cases and their results and the final decision on the status of the system.
- **Section 7 Attachments** is used to file the documentation of tester's training record, objective evidence, additional testing, Discrepancy Report, Problem Resolution Report, and Maintenance documentation.

Associates of Cape Cod, Inc. (ACC) aims to provide guidance, expertise and onsite assistance with the verification of the Lab Kinetics Incubating 8-well Tube Reader and BG Analytics® software. An Authorized Individual (representing the laboratory) shall identify whether the scope of this System Verification Protocol meets the local requirements, needs and expectations and may further modify this System Verification Plan. Section 2.4 Laboratory's System Verification Plan (Table 4) shall be used to formally document the plan specific for the laboratory where the system is to be permanently placed.

In an event of a discrepancy in Procedure, Expected Results or Observed Results within a certain Test Case, a Discrepancy Report shall be used to document the issue. The Discrepancy Report should include the following: reference to the Test Case, report number, description of the discrepancy, discrepancy investigation, description of resolution and category of resolution.

A failed Test Case may be re-executed following appropriate documentation of the failure on the Discrepancy Report. A pre-approved Problem Resolution Report should be used to document the procedure for re-execution of the Test Case. The Problem Resolution Report should include the following: reference to the Test Case, failure description, corrective actions, expected results and observed results.

Templates for Discrepancy Report, Problem Resolution Report, additional testing and maintenance are available electronically per request.

2.2 Description of the Components to Be Tested

There are three components that will be tested within the scope of this System Verification Protocol.

2.2.1 Lab Kinetics Incubating 8-well Tube Reader

The PKF08 instrument is an incubating absorbance tube reader equipped with eight (8) wells. Each well is individually read and timed, initiating data collection immediately upon insertion of a reaction tube. The PKF08 instrument is designed to equilibrate at and hold a temperature of 37°C ± 1°C during the 10-minute incubation step and during the assay runtime of 40 minutes. Digital Values are collected at two wavelengths: 405 nm (primary) and 495 nm (secondary) transmitted from the PKF08 instrument to a computer hosting BG Analytics® software. The PKF08 instrument is designed to accept tubes 12 mm in diameter. During incubation, 12x75 mm depyrogenated borosilicate glass may be used for sample preparation and pre-treatment. However, it is critical that the assay is performed in 12x65 mm flat-bottom tubes which are supplied as Fungitell STAT® Reagent.

Environmental requirements for operating PKF08 are described in Table 1. For more information, refer to Lab Kinetics Incubating 8-well Tube Reader User Manual (Incubating Kinetic Tube Reader User Manual) which is provided as a hard copy with the PKF08 instrument (or is available for download at www.acciusa.com).

Table 1. Environmental Requirements for the PKF08 Instrument

Environmental Requirements for PKF08	Description
Laboratory Conditions	Level and stable surface, away from equipment that may cause excessive vibration or electronic noise Avoid direct sunlight
Ambient Temperature	15°C – 30°C
Ambient Humidity	< 70%
Input Power	100-240 VAC @ 50/60 Hz
Connection to a power outlet	Power conditioner recommended Uninterruptable Power Supply (UPS) (optional)

2.2.2 BG Analytics® Software

The digital values transmitted by PKF08 are received by BG Analytics® software and converted to Optical Density values (OD). Data reduction involves calculation of rate (slope) from the kinetic data set Delta OD (405 – 495 nm) by fitting a linear regression to the range between 1900 and 2400 seconds.

The BG Analytics® software writes the collected data into an unshared local SQLite database referred to as BG Analytics database. The database provides search capabilities based on several criteria. For more information, refer to BG Analytics® User Manual G_1867.

The BG Analytics® software should be installed on a compatible host computer meeting the minimal requirements as described in Table 2:

Table 2: The Minimum System Requirements for the Computer Hosting BG Analytics® Software

System Requirement for Host Computer	Description
Operating System	Microsoft® Windows® 10 64 bit, version 1809 or newer
Physical Memory	Minimum: 4 GB Recommended: 8 GB
Hard Disk Space	Minimum: 10 GB Recommended: 15 GB and more
Communication Ports	At least one free USB port (or two (2) when using barcode scanner)

Additional requirements:

- A Microsoft® Windows user account
 - BG Analytics® software is installed on the host computer with SQLite database installed locally per user account:
 - A shared laboratory Microsoft® Windows user account may be utilized.
 - If multiple Microsoft® Windows user accounts are to be used, BGA has to be individually installed for each one of them.
- Connection to a barcode scanner (optional)
 - BGA is designed to be compatible with any barcode scanner that is configured in USB HID Points of Sale scanner mode. For example, Honeywell healthcare corded barcode scanners (e.g. Honeywell PN 1950HHD, Honeywell 1950HSR). Refer to the barcode scanner’s user manual for more information.
- Connection to a printer
- Anti-virus information
 - It is strongly recommended that an antivirus software with the most current update is installed and running on the computer hosting BG Analytics®. ACC recommends following local laboratory security policies.

2.2.3 Fungitell STAT® Assay

Within the BG Analytics® software, the slope of the sample is compared to the slope of the Standard yielding an Index value. Sample Index value is categorically interpreted as a Negative, Indeterminate, or Positive result according to the index value category ranges provided in Table 3. For further information, refer to Fungitell STAT® Instructions for Use (PN002603).

Table 3. Index Ranges as Described in Fungitell STAT® Instructions for Use

Fungitell STAT® Reportable Results	
Result	Index Value
Negative	≤ 0.74
Indeterminate	0.75 – 1.1
Positive	≥ 1.2

Note: The Fungitell STAT® assay is for in-vitro diagnostic use in the serum of patients. As such, it is recommended to perform the assay within a biological safety cabinet to increase operator’s safety while working with clinical

samples. This System Verification Protocol does not include the use of clinical samples, however, it is recommended that the Protocol be executed under environmental conditions that match that of intended use, thus within a biological safety cabinet.

2.3 Required Specifications

The required specifications for Lab Kinetics Incubating 8-well Tube Reader and BG Analytics® software are listed below:

- The PKF08 instrument must be calibrated at ACC prior to installation in the laboratory.
- The PKF08 instrument must be installed according to the manufacturer's requirements and environmental specifications. The PKF08 instrument must be shown to perform as determined based on the following performance data and their specifications:
 - Temperature – reported as a mean of temperatures measured over 5 minutes by the built-in NIST traceable temperature sensor.
 - Signal – mean well intensity (Digital Values, DVs) as measured over 5 minutes
 - Signal to Noise Ratio – Standard deviation of Digital Values over 5 minutes.
- The PKF08 instrument must be able to transmit data over time to BG Analytics® at the specified wavelengths, 405 nm and 495 nm, including incubation temperature.
- BG Analytics® must be installed at the laboratory according to the requirements.
- BG Analytics® must accept Fungitell STAT® Reagent, Fungitell STAT® Standard, LRW, APS, and patient sample identifiers when using a barcode scanner.
- The PKF08 instrument and BG Analytics® must collect, analyze and save test data in the embedded database after assay completion when used with Fungitell STAT® assay as an aid in clinical diagnostics of Invasive Fungal Infection.
- BG Analytics® must display patient test result on the screen at the completion of the test.
- BG Analytics® must display either a categorically negative result or invalid result when LRW is used as a negative control.
- BG Analytics® must display the kinetic trace of the sample when certain invalid Quality Conditions are identified.
- BG Analytics® must provide a printable and exportable report with one sample ID per page.
- BG Analytics® must provide capabilities for searching within the database by Standard lot number, Reagent lot number, Sample ID, and User ID.
- BG Analytics® must provide backup capability of the SQLite database.

2.4 Laboratory's System Verification Plan

This System Verification Protocol may be executed in full as written or, alternatively; an Authorized Individual (as recorded in Section 2.5.3 Personnel Log) may identify and record section(s) of this Protocol as Not Applicable (N/A) and/or define Additional Testing to meet the local requirements, needs and expectations. Table 4 should be used to record which section (if any) is N/A, initialed and dated.

Table 4. Laboratory System Verification Plan

Section No.	Section Description	Component Tested	Not Applicable? Initial/Date
3	IQ of PKF08	Lab Kinetics Incubating 8-well Tube Reader	<input type="checkbox"/> N/A _____
4	IQ of BGA	BG Analytics® software	<input type="checkbox"/> N/A _____
5	OQ of PKF08 and BGA	Lab Kinetics Incubating 8-well Tube Reader and BG Analytics® software	<input type="checkbox"/> N/A _____
7.3	Additional Testing	_____	<input type="checkbox"/> N/A _____

Section(s) recorded as N/A will not be executed and will not be taken into consideration when deliberating whether System Verification Protocol is a PASS in **Section 6 Final Verification Report**.

2.5 Responsibilities

The responsibilities are categorized as:

2.5.1 Vendor

This System Verification Protocol is designed to be executed by a trained individual representing the vendor. Contact information for the vendor supplying the PKF08 instrument, BG Analytics® software and Fungitell STAT® should be filled out in Table 5.

Table 5. Vendor Contact Information

Vendor Information	
Name	Associates of Cape Cod, Inc.
Address	124 Bernard E. Saint Jean Drive East Falmouth MA 02536 USA
Phone No.	001-508-540-3444
Technical Services Contact	e-mail: TechnicalServices@acciusa.com Phone No.: 001-888-848-3248
Local Authorized Vendor	Name: e-mail: Phone No.

2.5.2 Laboratory

This System Verification Protocol is expected to be reviewed and accepted by the Laboratory where the system will be permanently placed. Laboratory information should be filled out in Table 6.

Table 6. Laboratory Information

Laboratory Information	
Laboratory Name	
Company/Hospital Name	
Address	
Phone No.	
Primary contact	Name: e-mail: Phone No.

2.5.3 Personnel Log

Record the name and title of an Authorized Individual (representing the laboratory above) responsible for overseeing the placement of the PKF08 instrument and BG Analytics® software (including the execution of this Protocol):

Role: Authorized individual	
Name:	Title:
Signature:	Date:

Record the name and title of all personnel involved in the execution of this Protocol:

Role: Tester	
Name:	Title:
Signature:	Date:

Role: Reviewer	
Name:	Title:
Signature:	Date:

Role: _____	
Name:	Title:
Signature:	Date:

2.5.4 Documentation of Training on this System Verification Protocol

Document in **Section 7 Attachments** that individuals listed in Section 2.5.3 Personnel Log as Testers are trained to the content of this Protocol.

2.6 List of Supplies Required to Execute this System Verification Protocol

A list of supplies needed to execute this Protocol in full is provided in Table 7. All materials must be free of interfering glucans. Glassware must be dry-heat depyrogenated for at least 7 hours at a minimum of 235°C (or a validated equivalent) to be considered suitable for use.

Table 7. Required Supplies

Supplies	Vendor	ACC US Catalog Number*	Amount Needed	Storage Conditions
Lab Kinetics Incubating 8-well Tube Reader and BG Analytics®	ACC	PKF08-PKG	1	Ambient
Fungitell STAT® kit (10 vials of STAT Reagent + 5 vials of STAT Standard)	ACC	FT007	2 kits	2 – 8°C
Alkaline Pretreatment Solution (APS)	ACC	APS51-5	1 vial	2 – 30°C
250 µL pipette tips	ACC*	PPT25	1 pack	Ambient
1000 µL pipette tips	ACC*	PPT10	1 pack	Ambient
Long pipette tips 20 - 200 µL	ACC*	TPT50	1 pack	Ambient
12x75 mm depyrogenated borosilicate glass tubes	ACC	TB240-5	1 pack	Ambient
LAL Reagent Water (LRW)	ACC	W0051-10	1 bottle	2 – 30°C
Tube racks to fit tubes 12 mm in diameter	Any		2	Ambient
Vortex mixer	Any		1	Ambient
Parafilm® M	Any		1	Ambient
Adjustable pipette for volumes 100 - 1000 µL	Any		1	Ambient
Adjustable pipette for volumes 20 - 200 µL	Any		1	Ambient

*Or equivalent as available from regional authorized vendor

2.7 Procedure

Follow the procedure outlined below in the order described. Test Cases within each section are provided to generate objective evidence that the PKF08 instrument and BG Analytics® software meet the Required Specifications.

- Personnel executing or reviewing any section of this Protocol must complete Personnel Log in Section 2.5.3 Personnel Log.
- Personnel executing this Protocol must complete all Sections of this Protocol unless recorded as N/A in Table 4.
- Within each Section, an Authorized Individual may identify, record and appropriately justify if any Test Case(s) is N/A.
- Personnel executing this Protocol must execute all Test Cases within the applicable Section except those recorded as N/A.
- Personnel executing this Protocol must collect the objective evidence as defined in Procedure of each Test Case and document the Observed Results.
- Personnel executing this Protocol must print all objective evidence as defined in Expected Results (screen captures, reports etc.), label with the reference number and file in **Section 7 Attachments**.
- Personnel executing this Protocol must document PASS or FAIL status (except where N/A) for each Test Case.
- Personnel executing this Protocol must record any discrepancy from the Expected Results on Discrepancy

Report and must file the Report in **Section 7 Attachments**.

- Personnel executing this Protocol must follow a pre-approved Problem Resolution Report to resolve a problem and must file the Report in **Section 7 Attachments**.
- An Authorized Individual must review, sign and date each Test Case, including objective evidence, Discrepancy Report and Problem Resolution Report (if any). Discrepancy Report and Problem Resolution Report must be taken into account when making a decision regarding the status of the impacted Test Case.
- An Authorized Individual must identify and prepare a Test Case for additional testing (if any). Test Cases for additional testing must be filed in **Section 7 Attachments**.
- Personnel executing this Protocol must complete, sign and date Section 6.1 Final Verification Report.
- Two Authorized Individuals must review and approve each applicable section of this Protocol.
- Maintenance of the verified system (e.g. re-calibration of PKF08, database cleanup or upgrade of BGA software) may be tracked and filed in **Section 7 Attachments**.
- An Authorized Individual must file the completed System Verification Protocol in a location as indicated in Section 2.9 Location of the Completed System Verification Protocol.

2.8 Acceptance Criteria

- Each applicable Test Case shall PASS in order for the section of this Protocol to be considered conforming. A single Test Case that is recorded as FAIL indicates non-conformance of the entire section unless otherwise justified by an Authorized Individual.
- A Test Case that is recorded as FAIL cannot be re-executed without a documented Discrepancy Report and Problem Resolution Report which must be pre-approved by an Authorized Individual and attached to **Section 7 Attachments**.
- Each applicable Section of this Protocol must conform to the Required Specifications in order for the System Verification Protocol to PASS. The decision should be recorded in **Section 6 Final Verification Report**.

2.9 Location of the Completed System Verification Protocol

Upon completion and review, this System Verification Protocol will be filed in:

2.10 Review and Approval

This completed **Section 2**, identified as **System Verification Plan** of this System Verification Protocol adequately describes how to document that Lab Kinetics Incubating 8-well Tube Reader and BG Analytics® software meet the intended purpose and function.

Review and Approval	
_____	_____
Signature: Authorized Individual	Date
 Title	
_____	_____
Signature: Authorized Individual	Date
 Title	

3 Installation Qualification of Lab Kinetics Incubating 8-well Tube Reader

3.1 Calibration Documentation Test Case	
<input type="checkbox"/> N/A <i>Justification:</i> _____ <i>Initial/Date:</i> _____	
Purpose:	<i>The PKF08 instrument must be calibrated at ACC prior to installation in the laboratory.</i>
Test Procedure:	The PKF08 instrument is supplied with a Certificate of Calibration. This document provides evidence that the PKF08 instrument critical functions are calibrated within manufacturer’s specifications.
Expected Results:	A Certificate of Calibration is provided with the PKF08 instrument.
Observed Results:	A Certificate of Calibration is provided: <input type="checkbox"/> Yes, Date of Calibration: _____ <input type="checkbox"/> No
Discrepancy Report #:	
Pass or Fail:	
Performed By: (Sign/Date)	
Reviewed By: (Sign/Date)	

3.2 Setup of Lab Kinetics Incubating 8-well Tube Reader Test Case

□ N/A Justification: _____ Initial/Date: _____

Purpose:	<i>The PKF08 instrument must be installed according to the manufacturer's requirements and environmental specifications.</i>
Prerequisites:	The PKF08 instrument was received. TC 3.1 was successfully executed.
References:	PKF08 Incubating Kinetic Tube Reader User Manual
Test Procedure:	<ol style="list-style-type: none"> 1. Carefully open the PKF08 instrument packaging and transfer the PKF08 instrument to a clean, flat surface. 2. Visually inspect the exterior of the PKF08 instrument for any signs of damage e.g. scratches and record any observations in Observed Results. 3. Remove all the remaining components (power cord, USB communication cable, power conditioner and dust cover) from the box and inspect them for any signs of damage. Record any observations in Observed Results. 4. If any materials are missing or damaged, contact Technical Services at TechnicalServices@acciusa.com.
Expected Results:	<ul style="list-style-type: none"> • The PKF08 instrument is present and undamaged. • All the remaining components are present and undamaged.
Observed Results:	<ul style="list-style-type: none"> • The PKF08 instrument is present and undamaged: <input type="checkbox"/>Yes <input type="checkbox"/>No, _____ • All the remaining components are present and undamaged: <input type="checkbox"/>Yes <input type="checkbox"/>No, _____
Discrepancy Report #:	
Pass or Fail:	
Performed By: (Sign/Date)	
Reviewed By: (Sign/Date)	

3.3 The Installation Lab Kinetics Incubating 8-well Tube Reader Test Case

N/A Justification: _____ Initial/Date: _____

Purpose:	The PKF08 instrument must be installed according to the manufacturer's requirements and environmental specifications.
Prerequisites:	Thermometer identification: Model: _____ Serial Number: _____ Cal Due: _____ Hygrometer identification: Model: _____ Serial Number: _____ Cal Due: _____
References:	PKF08 Incubating Kinetic Tube Reader User Manual
Test Procedure:	<ol style="list-style-type: none"> Record the laboratory environmental conditions in Observed Results. Confirm that the environmental conditions meet the requirements in Observed Results. Electricity is not measured, only recorded based on the type of grid. If the environmental requirements are met, connect the PKF08 instrument to a grounded wall outlet via power conditioner or UPS. Record the PKF08 instrument information in Observed Results.
Expected Results:	<ul style="list-style-type: none"> The environmental conditions are documented and meet the requirements. The PKF08 Instrument information is documented. The PKF08 instrument is installed.
Observed Results:	<ul style="list-style-type: none"> Environmental conditions: <ul style="list-style-type: none"> Ambient temperature: _____ °C (required 15 - 30°C) Ambient humidity: _____ % (required < 70%) Electricity: _____ VAC (required 100-240VAC @ 50/60 Hz) Environmental conditions meet the requirements: <input type="checkbox"/>Yes <input type="checkbox"/>No, _____
Observed Results:	<ul style="list-style-type: none"> PKF08 Instrument information: <ul style="list-style-type: none"> Serial Number: _____ Connection via a power conditioner: <input type="checkbox"/>Yes <input type="checkbox"/>No Make/Model: _____ Connected via UPS (optional): <input type="checkbox"/>Yes <input type="checkbox"/>No Make/Model: _____ The PKF08 Instrument is installed: <input type="checkbox"/>Yes <input type="checkbox"/>No
Discrepancy Report #:	
Pass or Fail:	
Performed By: (Sign/Date)	
Reviewed By: (Sign/Date)	

3.4 Evaluation of Lab Kinetics Incubating 8-well Tube Reader Test Case

N/A Justification: _____ Initial/Date: _____

Purpose:	<i>The PKF08 instrument must be installed according to manufacturer's requirements and environmental specifications.</i>
Prerequisites:	TC 3.3 was completed. Set up eight 12x75 mm depyrogenated borosilicate glasstubes.
References:	Incubating Kinetic Tube Reader User Manual
Test Procedure:	<ol style="list-style-type: none"> 1. Power on PKF08. 2. Allow PKF08 to go through initialization. 3. Insert 12x75 mm tubes in all eight wells. 4. Observe the performance of the PKF08 instrument as explained in Expected Results. Document in Observed Results.
Expected Results:	<ul style="list-style-type: none"> • After turning PKF08 on – no tubes inserted: <ul style="list-style-type: none"> ○ The LCD screen is on ○ The LCD screen displays serial number and a wavelength ○ All empty well LEDs are red • After inserting 12x75 mm tubes in all eight wells: <ul style="list-style-type: none"> ○ All tubes can be inserted completely ○ All the well LEDs turn green
Observed Results:	<ul style="list-style-type: none"> • After turning PKF08 on – no tubes inserted: <ul style="list-style-type: none"> ○ The LCD screen on <input type="checkbox"/>Yes <input type="checkbox"/>No ○ The LCD screen displays serial number, wavelength <input type="checkbox"/>Yes <input type="checkbox"/>No ○ All empty well LEDs red <input type="checkbox"/>Yes <input type="checkbox"/>No • After inserting 12x75 mm tubes in all eight wells: <ul style="list-style-type: none"> ○ All tubes can be inserted completely <input type="checkbox"/>Yes <input type="checkbox"/>No ○ All the well LEDs turn green <input type="checkbox"/>Yes <input type="checkbox"/>No
Discrepancy Report #:	
Pass or Fail:	
Performed By: (Sign/Date)	
Reviewed By: (Sign/Date)	

3.5 Verification of the Performance of Lab Kinetics Incubating 8-well Tube Reader Test Case

N/A Justification: _____ Initial/Date: _____

Purpose:	<p>The PKF08 instrument must be shown to perform as determined based on the following performance data and their specifications:</p> <ul style="list-style-type: none"> • Temperature – reported as a mean of temperatures measured over 5 minutes by the built-in NIST traceable temperature sensor • Signal – mean well intensity (Digital Values, DVs) as measured over 5 minutes • Signal to Noise Ratio – Standard deviation of Digital Values over 5 minutes
Prerequisites:	<p>PKF08 has been on for at least 20 minutes.</p> <p>TC 3.4 was completed.</p> <p>All tubes are removed from PKF08.</p> <p>External computer with PKF08 Calibration Tool Version number: _____</p>
Test Procedure:	<ol style="list-style-type: none"> 1. Using the USB communication cable, connect the PKF08 instrument to the external computer hosting PKF08 Calibration Tool. 2. Launch PKF08 Calibration Tool. 3. On the home screen, select the PKF08 serial number from the Instrument dropdown menu. 4. Click Auto Calibrate. 5. Allow the PKF08 Calibration Tool to proceed with the calibration process. 6. Once completed, click Print and label as 3.5_6. 7. Evaluate the As-Found and As-Left data listed under Performance Results per specifications stated in Expected Results. 8. Document in Observed Results. 9. Close the PKF08 Calibration Tool.
Expected Results:	<ul style="list-style-type: none"> • As shown in TC 3.5_6, under Performance Results: <ul style="list-style-type: none"> ○ Active 405nm bank set is identified ○ Reported Mean Temperature: $37 \pm 1^\circ\text{C}$ ○ Optical Intensity for active 405nm bank for each well number: $\geq 36,000$ ○ Optical Intensity for bank 495nm for each well number: $\geq 36,000$ ○ Signal to Noise Ratio for active 405nm bank: > 261 ○ Signal to Noise Ratio for 495nm bank: > 261
Observed Results:	<ul style="list-style-type: none"> • As shown in TC 3.5_6, As Found Performance Results: <ul style="list-style-type: none"> ○ Active 405nm bank: _____ ○ Mean Temperature: $37 \pm 1^\circ\text{C}$: <input type="checkbox"/>Yes <input type="checkbox"/>No ○ Optical Intensity for active 405nm bank for each well number $\geq 36,000$: <input type="checkbox"/>Yes <input type="checkbox"/>No ○ Optical Intensity for bank 495nm for each well number $\geq 36,000$: <input type="checkbox"/>Yes <input type="checkbox"/>No ○ Signal to Noise Ratio for active 405nm bank > 261: <input type="checkbox"/>Yes <input type="checkbox"/>No ○ Signal to Noise Ratio for 495nm bank: > 261: <input type="checkbox"/>Yes <input type="checkbox"/>No • As-Left Performance Results: <ul style="list-style-type: none"> ○ Active 405nm bank: _____ ○ Mean Temperature: $37 \pm 1^\circ\text{C}$: <input type="checkbox"/>Yes <input type="checkbox"/>No ○ Optical Intensity for active 405nm bank for each well number $\geq 36,000$: <input type="checkbox"/>Yes <input type="checkbox"/>No ○ Optical Intensity for bank 495nm for each well number $\geq 36,000$: <input type="checkbox"/>Yes <input type="checkbox"/>No ○ Signal to Noise Ratio for active 405nm bank > 261: <input type="checkbox"/>Yes <input type="checkbox"/>No ○ Signal to Noise Ratio for 495nm bank: > 261: <input type="checkbox"/>Yes <input type="checkbox"/>No
Discrepancy Report #:	

Pass or Fail:	
Performed By: (Sign/Date)	
Reviewed By: (Sign/Date)	

3.6 Review and Approval

This completed **Section 3**, identified as **Installation Qualification of Lab Kinetics Incubating 8-well Tube Reader**, documents that the PKF08 instrument has passed all testing of the specified processes for which it was intended.

Review and Approval	
_____	_____
Signature: Authorized Individual	Date
 Title	
_____	_____
Signature: Authorized Individual	Date
 Title	

4 Installation Qualification of BG Analytics® Software

4.1 Installation of BG Analytics® Software Test Case	
<input type="checkbox"/> N/A Justification: _____ Initial/Date: _____	
Purpose:	BG Analytics® software must be installed at the laboratory according to the manufacturer's requirements.
Prerequisites:	<p>Computer meeting minimum system requirements (Win10 64-bit, version 1809 or newer) with at least one available USB port ready for install.</p> <p>A dedicated local Windows® user account.</p> <p>Download BG Analytics® software from ACC software portal https://portal.acciusa.com following instructions within the BG Analytics® User Manual (G_1867) at Section 1.3 for registration steps and Section 2.5 for Installation steps.</p>
References:	BG Analytics® User Manual (G_1867) ACC software portal https://portal.acciusa.com
Test Procedure:	<ol style="list-style-type: none"> 1. In Observed Results, confirm that the computer specifications meet the minimum requirements. 2. In Observed Results, record the computer ID, dedicated User ID and BG Analytics® software version. 3. Install the BG Analytics® software onto the host computer under the dedicated local Windows® user ID. 4. Upon first-time launch, review and Accept the BG Analytics® Software End User License Agreement to proceed to the Home screen. 5. Take a screenshot of the BG Analytics® Home screen. 6. Save the screenshot as TC 4.1_1. 7. Verify that BG Analytics® Home page displays Start Test and View Results. 8. Close BG Analytics®. 9. On the computer, navigate to Start and right-click over BG Analytics®. Click More, then Pin to taskbar to create an icon on the taskbar.
Expected Results:	<ul style="list-style-type: none"> • Computer meets minimum system requirements. • BG Analytics® software was successfully installed. • As shown in TC 4.1_1, BG Analytics® Home page displays Start Test and View Results.
Observed Results:	<ul style="list-style-type: none"> • Computer meets minimum system requirements: <input type="checkbox"/>Yes <input type="checkbox"/>No • Computer and software information: <ul style="list-style-type: none"> ○ Computer ID: _____ ○ User ID on the host computer: _____ ○ BG Analytics® software version: _____ • BG Analytics® software was successfully installed: <input type="checkbox"/>Yes <input type="checkbox"/>No • As shown in TC 4.1_1, BG Analytics® Home page displays Start Test and View Results: <input type="checkbox"/>Yes <input type="checkbox"/>No
Discrepancy Report #:	

Pass or Fail:	
Performed By: (Sign/Date)	
Reviewed By: (Sign/Date)	

4.2 Installation of Barcode Scanner Test Case

N/A Justification: All information will be entered manually only (keyboard entry) Initial/Date: _____

Purpose:	BG Analytics [®] must accept Fungitell STAT [®] Reagent, Fungitell STAT [®] Standard, and patient sample identifiers when using a barcode scanner.
Prerequisites:	A configured a barcode scanner meeting vendor's recommendation. BG Analytics [®] is installed and closed.
References:	BG Analytics [®] User Manual (G_1867) Barcode scanner User Manual
Test Procedure:	<ol style="list-style-type: none"> 1. In Observed Results, record the Barcode scanner description. 2. Install a configured scanner on the host computer by following the manufacturer's installation procedure. 3. Launch BG Analytics[®]. 4. Click Start Test. 5. Once on the Test Setup screen, scan available barcodes (if any). 6. Take a screenshot of the filled fields on Test Setup screen. 7. Save the screenshot as TC 4.2_1. 8. Verify that the information on barcoded items was appropriately filled in BGA.
Expected Results:	<ul style="list-style-type: none"> • Barcode scanner meets vendor's recommendations. • Barcode scanner was successfully installed. • As shown in TC 4.2_1, BG Analytics[®] Test Setup screen appropriately fills all the barcoded information.
Observed Results:	<ul style="list-style-type: none"> • Barcode scanner description: _____ • Barcode scanner meets vendor's recommendations: <input type="checkbox"/>Yes <input type="checkbox"/>No • Barcode scanner was successfully installed: <input type="checkbox"/>Yes <input type="checkbox"/>No • As shown in TC 4.2_1, BG Analytics[®] Test Setup screen appropriately fills all the barcoded information: <input type="checkbox"/>Yes <input type="checkbox"/>No
Discrepancy Report #:	
Pass or Fail:	
Performed By: (Sign/Date)	
Reviewed By: (Sign/Date)	

4.3 Review and Approval

This completed **Section 4**, identified as **Installation Qualification of BG Analytics® Software**, documents that the software is adequately installed and has passed all testing of the specified processes for which it was intended.

Review and Approval	
_____ Signature: Authorized Individual	_____ Date
 Title	
_____ Signature: Authorized Individual	_____ Date
 Title	

5 Operational Qualification of Lab Kinetics Incubating 8-well Tube Reader and BG Analytics® Software

5.1 Verification of Data Transmission Test Case	
<input type="checkbox"/> N/A Justification: _____ Initial/Date: _____	
Purpose:	The PKF08 instrument must be able to transmit data over time to BG Analytics® software at 405 nm and 495 nm, including the incubation temperature.
Prerequisites:	IQ of PKF08 and IQ of BGA were completed. PKF08 has been on for at least 20 minutes. All tubes are removed from PKF08.
References:	BG Analytics® User Manual (G_1867)
Test Procedure:	<ol style="list-style-type: none"> 1. Launch BG Analytics®. 2. Click on Start Test. 3. BGA displays Verifying Instrument screen and goes through a minimum of 30-second self-test. 4. Take a screenshot of Verifying Instrument screen. 5. Save the screenshot as TC 5.1_1. 6. Verify that BGA displays all the parameters as listed in Expected Results. 7. Upon completion of the self-test, BGA switches to Test Setup screen. 8. Take a screenshot of Test Setup screen. 9. Save the screenshot as TC 5.1_2. 10. Verify that the transmitted temperature is 37°C ± 1°C.
Expected Results:	<ul style="list-style-type: none"> • As shown in TC 5.1_1, BG Analytics® Verifying Instrument screen displays: <ul style="list-style-type: none"> ○ Self-test in progress... ○ PKF08 serial number ○ Transmitted temperature ○ Status: Connected • As shown in TC 5.1_2, following self-test, BGA proceeded to Test Setup screen. • As shown in TC 5.1_2, the transmitted temperature is 37°C ± 1°C.
Observed Results:	<ul style="list-style-type: none"> • As shown in TC 5.1_1, BG Analytics® Verifying Instrument screen displays: <ul style="list-style-type: none"> ○ Self-test in progress... <input type="checkbox"/>Yes <input type="checkbox"/>No ○ PKF08 serial number: <input type="checkbox"/>Yes <input type="checkbox"/>No ○ Transmitted temperature: <input type="checkbox"/>Yes <input type="checkbox"/>No ○ Status: Connected <input type="checkbox"/>Yes <input type="checkbox"/>No • As shown in TC 5.1_2, following self-test, BGA proceeded to Test Setup screen: <input type="checkbox"/>Yes <input type="checkbox"/>No • As shown in TC 5.1_2, the transmitted temperature is 37°C ± 1°C: <input type="checkbox"/>Yes <input type="checkbox"/>No
Discrepancy Report #:	
Pass or Fail:	

Performed By: (Sign/Date)	
Reviewed By: (Sign/Date)	

5.2 Verification of Collecting, Saving, Analysis and Test Result Delivery Test Case

□ N/A Justification: _____ Initial/Date: _____

Purpose:	<p><i>The PKF08 instrument and BG Analytics® must collect, analyze and save test data in the embedded database after assay completion when used with Fungitell STAT® assay as an aid in clinical diagnostics of Invasive Fungal Infection.</i></p> <p><i>BG Analytics® must display patient test results on the screen at the completion of the test.</i></p> <p><i>BG Analytics® must provide a printable and exportable report with one sample ID (patient result) per page.</i></p>
Prerequisites:	<p>IQ of PKF08 and IQ of BGA were completed.</p>
References:	<p>BG Analytics® User Manual (G_1867) Fungitell STAT® Instructions for Use (PN002603)</p>
Test Procedure:	<ol style="list-style-type: none"> 1. Launch BG Analytics®. 2. Click Start Test. 3. Wait for Test Setup screen. 4. Type in User ID. 5. Use the installed barcode scanner or type in lot number and expiry information for each field (Standard lot, Reagent lot, APS lot, Water lot). 6. Type in Sample ID for all seven (7) samples as "OQ1", "OQ2" etc. 7. Take a screenshot of Test Setup screen. 8. Save the screenshot as TC 5.2_1. 9. Verify that all data entry is correctly displayed on Test Setup screen. 10. Click Start to proceed to Incubating screen. 11. Prepare two (2) Fungitell STAT® STD (STAT STD) tubes: <ol style="list-style-type: none"> a. Reconstitute each with the specific volume of LRW as per label, vortex for 15 seconds and cover. b. To each tube, add the specific volume of APS as per label, vortex for 15 seconds and cover. 12. On Incubating screen, insert both STAT STD tubes in any well of PKF08 for a 10-minute incubation. 13. Take a screenshot of the Incubating screen. 14. Save the screenshot as TC 5.2_2. 15. Verify that the status of two wells is "Incubating" and both timers count down from 10:00 minutes. 16. During incubation, reconstitute eight (8) Fungitell STAT® RGT (STAT RGT) tubes with 300 µL of LRW and vortex mix each tube for no more than 5 seconds. 17. When the well status of both tubes changes to "Done Incubating", remove both from PKF08 and pool by pipetting the entire volume from one tube to the other. 18. Vortex the pooled STAT STD tube for 15 seconds. 19. Transfer 75 µL from the STAT STD pool to each of the eight STAT RGT.

	<ol style="list-style-type: none"> 20. Vortex each STAT RGT for no more than 5 seconds and cover. 21. In BGA, when prompted to proceed to data collection, click Yes. 22. On Collecting Data screen, insert each STAT RGT tube individually into PKF08 to start a 40-minute data collection. 23. Take a screenshot of Collecting Data screen. 24. Save the screenshot as TC 5.2_3. 25. Verify that the status of all wells is “Collecting” and all timers count down from 40:00 minutes. 26. Allow the test to run to completion. 27. When BGA displays “The test has finished”, click View Results. 28. Take a screenshot of the BG Analytics® Test Result screen. 29. Save the screenshot as TC 5.2_4. 30. Verify that Test Result screen displays a header containing the test information and test results for samples OQ1 and OQ2. 31. Click Print to print the entire report consisting of 7 pages. 32. Label each page from TC 5.2_5 to TC 5.2_11. 33. Verify that each page of the report displays the parameters as defined in Expected Results. 34. Click Export to export the report as BG Analytics file. Select a location of the export on the desktop and click Save. 35. Take a screenshot of the desktop. 36. Save the screenshot as TC 5.2_12. 37. Verify that the BG Analytics file was successfully exported. 38. Open the exported file and print the exported reports. 39. Label the reports as TC 5.2_13 to TC 5.2_19. 40. Verify that reports TC 5.2_13 to TC 5.2_19 match reports TC 5.2_5 to TC 5.2_11. 41. Close BG Analytics®.
<p>Expected Results:</p>	<ul style="list-style-type: none"> • As shown on TC 5.2_1, Test Setup screen displays all data entry correctly. • As shown on TC 5.2_2, the status of two wells is “Incubating” and both timers count down from 10:00 minutes. • As shown on TC 5.2_3, the status of all wells is “Collecting” and all timers count down from 40:00 minutes. • As shown on TC 5.2_4, the Test Result screen displays a header containing the test information and test results for samples OQ1 and OQ2. • As shown on TC 5.2_5 – TC 5.2_11, each page of the report displays the following parameters: <ul style="list-style-type: none"> ○ Header with test information ○ Sample ID ○ Sample section: <ul style="list-style-type: none"> ▪ QC Status: Valid – In Range ▪ Index: within the range of 0.75 – 1.2 ▪ Sample category: Indeterminate or Positive • As shown on TC 5.2_12, the report was exported as BG Analytics file. • As shown on TC 5.2_13 to TC 5.2_19, the exported reports match reports TC 5.2_5 to TC 5.2_11.

Observed Results:	<ul style="list-style-type: none"> • As shown on TC 5.2_1, Test Setup screen displays all data entry correctly: <input type="checkbox"/>Yes <input type="checkbox"/>No • As shown on TC 5.2_2, the status of two wells is “Incubating” and both timers count down from 10:00 minutes: <input type="checkbox"/>Yes <input type="checkbox"/>No • As shown on TC 5.2_3, the status of all wells is “Collecting” and all timers count down from 40:00 minutes: <input type="checkbox"/>Yes <input type="checkbox"/>No • As shown on TC 5.2_4, the Test Result screen displays a header containing the test information and test results for samples OQ1 and OQ2: <input type="checkbox"/>Yes <input type="checkbox"/>No • As shown on TC 5.2_5 – TC 5.2_11, each page of the report displays the following parameters: <ul style="list-style-type: none"> ○ Header with test information: <input type="checkbox"/>Yes <input type="checkbox"/>No ○ Sample ID: <input type="checkbox"/>Yes <input type="checkbox"/>No ○ Sample section: <ul style="list-style-type: none"> ▪ QC Status: Valid – In Range <input type="checkbox"/>Yes <input type="checkbox"/>No ▪ Index: within the range of 0.75 – 1.2 <input type="checkbox"/>Yes <input type="checkbox"/>No ▪ Sample category: Indeterminate or Positive <input type="checkbox"/>Yes <input type="checkbox"/>No • The Report was exported to BG Analytics file as shown in TC 5.2_12: <input type="checkbox"/>Yes <input type="checkbox"/>No • As shown on TC 5.2_13 to TC 5.2_19, the exported reports match reports TC 5.2_5 to TC 5.2_11. <input type="checkbox"/>Yes <input type="checkbox"/>No
Discrepancy Report #:	
Pass or Fail:	
Performed By: (Sign/Date)	
Reviewed By: (Sign/Date)	

5.3 Verification of BG Analytics® Reporting Test Results Test Case

□ N/A Justification: _____ Initial/Date: _____

Purpose:	<i>BG Analytics® must display either a categorically negative result or invalid result when LRW is used as a negative control.</i> <i>BG Analytics® must display kinetic trace of the sample when certain invalid QCs are identified.</i>
Prerequisites:	IQ of PKF08 and IQ of BG Analytics® were completed.
References:	BG Analytics® User Manual (G_1867) Fungitell STAT® Instructions for Use (PN002603)
Test Procedure:	<ol style="list-style-type: none"> 1. Launch BG Analytics®. 2. Click Start Test. 3. Wait for Test Setup screen. 4. Type in User ID. 5. Use the installed barcode scanner or type in lot number and expiry information for each field (Standard lot, Reagent lot, APS lot, Water lot). 6. Type in Sample IDs for Sample 1, 2 and 3 as LRW1, LRW2, LRW3. 7. Type in Sample IDs for Sample 4, 5 and 6 as Non recon 1, Non recon 2, Non recon 3. 8. Under Notes, type in the following text: "OQ TC 5.3" 9. Click Start to proceed to Incubating screen. 10. Prepare one STAT STD tube: <ul style="list-style-type: none"> • Reconstitute STAT STD with the specific volume of LRW as per label, vortex for 15 seconds and cover. • To STAT STD, add the specific volume of APS as per label, vortex for 15 seconds and cover. 11. Preparation of Sample 1, 2 and 3: <ul style="list-style-type: none"> • Transfer 50 µL of LRW into three empty 12x75 mm tubes. • Add 200 µL of APS to each. • Vortex mix for 15 seconds and cover. 12. On Incubating screen, insert STAT STD and Sample 1, 2 and 3 into the designated wells of PKF08 for a 10-minute incubation. 13. During incubation, reconstitute four (4) STAT RGT tubes with 300 µL of LRW and vortex mix each tube for no more than 5 seconds. Acquire an additional three (3) STAT RGT tubes but do <u>not</u> reconstitute them (for a total of four reconstituted and three non-reconstituted STAT RGT tubes). 14. When the well status changes to "Done Incubating", remove all tubes from PKF08 and vortex each tube for 5 seconds. 15. Transfer 75 µL from the STAT STD to a reconstituted STAT RGT tube. 16. Transfer 75 µL from each of the Sample 1, Sample 2 and Sample 3 tubes to corresponding reconstituted STAT RGT tubes. 17. Transfer 75 µL from the LRW to each of the three non-reconstituted STAT RGT tubes. 18. Vortex the first four (reconstituted)RGT tubes for no more than 5 seconds and cover. Do not vortex the non-reconstituted RGT tubes, only cover. 19. In BGA, when prompted to proceed to collection data, click Yes. 20. On Collecting Data screen, insert each STAT RGT tube individually into PKF08 to start a 40-minute data collection. 21. When BGA displays "The test has finished", click View Results. 22. Click Print to print the generated report. 23. Label the reports as TC 5.3_1 to TC 5.3_6. 24. Verify that the reports for Sample ID: LRW1, LRW2 and LRW3 display parameters as defined in Expected Results.

	<p>25. Verify that the reports for Sample ID: Non recon 1, Non recon 2, Non recon 3 display parameters as defined in Expected Results.</p> <p>26. Verify that the report displays the entered text in the header under Notes: "OQ TC 5.3".</p> <p>27. Close BGA software.</p>
<p>Expected Results:</p>	<ul style="list-style-type: none"> ● As shown in TC 5.3_1 to TC 5.3_6, the reports display the entered text in the header under Notes: "OQ TC 5.3". ● As shown in TC 5.3_1, TC 5.3_2 and TC 5.3_3, the reports for Sample ID: LRW1, LRW2, LRW3 display one of the listed outputs below: <ul style="list-style-type: none"> ○ Output 1 <ul style="list-style-type: none"> ○ Sample section: <ul style="list-style-type: none"> ▪ QC Status: Valid – Below Range ▪ Index: Index Not Calculated ▪ Sample Category: Negative ○ Output 2 <ul style="list-style-type: none"> ○ Sample section: <ul style="list-style-type: none"> ▪ QC Status: Invalid – Not Above 0 at 500 ▪ Index: Index Not Calculated ▪ Sample Category: Not Reportable ○ A plot of sample kinetic plot as (Delta OD (405 – 495 nm) vs. Time (s)) ○ Y-intercept, slope and R values determined between 1900 and 2400 seconds ● As shown in TC 5.3_4, TC 5.3_5 and TC 5.3_6, the reports for Sample ID: Non recon 1, Non recon 2, Non recon 3 display: <ul style="list-style-type: none"> ○ Sample section: <ul style="list-style-type: none"> ▪ QC Status: Invalid – Not Above 0 at 500 or Invalid-End OD ▪ Index: Index Not Calculated ▪ Sample Category: Not Reportable ○ A plot of sample kinetic plot as (Delta OD (405 – 495 nm) vs. Time (s)) ○ Y-intercept, slope and R values determined between 1900 and 2400
<p>Observed Results</p>	<ul style="list-style-type: none"> ● As shown in TC 5.3_1 to TC 5.3_6, the reports display the entered text in the header under Notes: "OQ TC 5.3": <input type="checkbox"/>Yes <input type="checkbox"/>No ● As shown in TC 5.3_1, TC 5.3_2 and TC 5.3_3, the reports for Sample ID: LRW1, LRW2, LRW3 display one of the listed outputs below: <ul style="list-style-type: none"> ○ Output 1 <ul style="list-style-type: none"> ○ Sample section: <ul style="list-style-type: none"> ▪ QC Status: Valid – Below Range <input type="checkbox"/>Yes <input type="checkbox"/>No ▪ Index: Index Not Calculated <input type="checkbox"/>Yes <input type="checkbox"/>No ▪ Sample Category: Negative <input type="checkbox"/>Yes <input type="checkbox"/>No ○ Output 2 <ul style="list-style-type: none"> ○ Sample section: <ul style="list-style-type: none"> ▪ QC Status: Invalid – Not Above 0 at 500 <input type="checkbox"/>Yes <input type="checkbox"/>No ▪ Index: Index Not Calculated <input type="checkbox"/>Yes <input type="checkbox"/>No ▪ Sample Category: Not Reportable <input type="checkbox"/>Yes <input type="checkbox"/>No ○ A plot of sample kinetic plot as (Delta OD (405 – 495 nm) vs. Time (s)): <input type="checkbox"/>Yes <input type="checkbox"/>No ○ Y-intercept, slope and R values determined between 1900 and 2400 seconds: <input type="checkbox"/>Yes <input type="checkbox"/>No ● As shown in TC 5.3_4, TC 5.3_5 and TC 5.3_6, the reports for Sample ID: Non recon 1, Non recon 2, Non recon 3 display: <ul style="list-style-type: none"> ○ Sample section: <ul style="list-style-type: none"> ▪ QC Status: Invalid – Not Above 0 at 500 or Invalid-End OD <input type="checkbox"/>Yes <input type="checkbox"/>No ▪ Index: Index Not Calculated <input type="checkbox"/>Yes <input type="checkbox"/>No ▪ Sample Category: Not Reportable <input type="checkbox"/>Yes <input type="checkbox"/>No ○ A plot of sample kinetic plot as (Delta OD (405 – 495 nm) vs. Time (s)) <input type="checkbox"/>Yes <input type="checkbox"/>No ● Y-intercept, slope and R values determined between 1900 and 2400 <input type="checkbox"/>Yes <input type="checkbox"/>No

Discrepancy Report #:	
Pass or Fail:	
Performed By: (Sign/Date)	
Reviewed By: (Sign/Date)	

5.4 Verification of Data Storage and Searching Capabilities Test Case

□ N/A Justification: _____ Initial/Date: _____

Purpose:	BG Analytics® must provide capabilities for searching within the database by Standard lot number, Reagent lot number, Sample ID and User ID.
Prerequisites:	IQ of PKF08 and IQ of BG Analytics® completed. TC 5.3 was completed.
References:	BG Analytics® User Manual (G_1867)
Test Procedure:	<ol style="list-style-type: none"> 1. Launch BG Analytics®. 2. Click View Results. 3. Click into Search box to locate the record by Sample ID. Type in “LRW1” which is the Sample ID. 4. Click Find to display the search result. 5. Take a screenshot of the Test History screen. 6. Label the screenshot as TC 5.4_1. 7. Verify that only result for sample “LRW1” is displayed. 8. Double-click over sample “LRW1” line and click Print to print the generated report. 9. Label the report as TC 5.4_2. 10. Verify that the same test report is generated as in TC 5.3_1. 11. Close BG Analytics®.
Expected Results:	<ul style="list-style-type: none"> ● As shown on TC 5.4_1, BGA allows searching by Sample ID. ● As shown on TC 5.4_2, after re-opening, the report for Sample “LRW1” is identical to TC 5.3_1.
Observed Results:	<ul style="list-style-type: none"> ● As shown on TC 5.4_1, BGA allows searching by Sample ID: <input type="checkbox"/>Yes <input type="checkbox"/>No ● As shown on TC 5.4_2, after re-opening, the report for Sample “LRW1” is identical to TC 5.3_1: <input type="checkbox"/>Yes <input type="checkbox"/>No
Discrepancy Report #:	
Pass or Fail:	
Performed By: (Sign/Date)	
Reviewed By: (Sign/Date)	

5.5 Verification of Database Backup Capability Test Case

N/A *Justification:* _____ *Initial/Date:* _____

Purpose:	<i>BG Analytics® must provide a capability for a backup of the SQLite database.</i>
Prerequisites:	IQ of PKF08 and IQ of BG Analytics® were completed.
References:	BG Analytics® User Manual (G_1867)
Test Procedure:	<ol style="list-style-type: none"> 1. Launch BG Analytics®. 2. Click Backup. 3. Select a location on the desktop of the host computer to save the copy of the database. 4. Click Save under the default file name in the format bgabackup-YEAR-MONTH-DAY as type: BGA database. 5. Click OK to confirm Backup Complete. 6. Take a screenshot of the desktop. 7. Save the screenshot as TC 5.5_1. 8. Verify that a file named bgabackup-YEAR-MONTH-DAY is displayed. 9. Close BG Analytics®.
Expected Results:	<ul style="list-style-type: none"> ● As shown on TC 5.5_1, a file named bgabackup-YEAR-MONTH-DAY is displayed.
Observed Results:	<ul style="list-style-type: none"> ● As shown on TC 5.5_1, a file named bgabackup-YEAR-MONTH-DAY is displayed: <input type="checkbox"/>Yes <input type="checkbox"/>No
Discrepancy Report #:	
Pass or Fail:	
Performed By: (Sign/Date)	
Reviewed By: (Sign/Date)	

5.6 Review and Approval

This completed **Section 5**, identified as **Operational Qualification of Lab Kinetics Incubating 8-well Tube Reader and BG Analytics® Software** documents that the system has passed all specified testing and will adequately perform when used for the purposes for which it was intended.

Review and Approval	
_____	_____
Signature: Authorized Individual	Date

Title	
_____	_____
Signature: Authorized Individual	Date

Title	

6 Final Verification Report

6.1 Final Verification Report	
Purpose:	To provide an overview of Test Case results
Section 3 Review:	<div style="text-align: right;">Section N/A <input type="checkbox"/></div> TC 3.1 Pass <input type="checkbox"/> Fail <input type="checkbox"/> N/A <input type="checkbox"/> TC 3.2 Pass <input type="checkbox"/> Fail <input type="checkbox"/> N/A <input type="checkbox"/> TC 3.3 Pass <input type="checkbox"/> Fail <input type="checkbox"/> N/A <input type="checkbox"/> TC 3.4 Pass <input type="checkbox"/> Fail <input type="checkbox"/> N/A <input type="checkbox"/> TC 3.5 Pass <input type="checkbox"/> Fail <input type="checkbox"/> N/A <input type="checkbox"/> Notes: Section Conforms to Required Specifications: YES <input type="checkbox"/> NO <input type="checkbox"/>
Section 4 Review:	<div style="text-align: right;">Section N/A <input type="checkbox"/></div> TC 4.1 Pass <input type="checkbox"/> Fail <input type="checkbox"/> N/A <input type="checkbox"/> TC 4.2 Pass <input type="checkbox"/> Fail <input type="checkbox"/> N/A <input type="checkbox"/> Notes: Section Conforms to Required Specifications: YES <input type="checkbox"/> NO <input type="checkbox"/>
Section 5 Review:	<div style="text-align: right;">Section N/A <input type="checkbox"/></div> TC 5.1 Pass <input type="checkbox"/> Fail <input type="checkbox"/> N/A <input type="checkbox"/> TC 5.2 Pass <input type="checkbox"/> Fail <input type="checkbox"/> N/A <input type="checkbox"/> TC 5.3 Pass <input type="checkbox"/> Fail <input type="checkbox"/> N/A <input type="checkbox"/> TC 5.4 Pass <input type="checkbox"/> Fail <input type="checkbox"/> N/A <input type="checkbox"/> TC 5.5 Pass <input type="checkbox"/> Fail <input type="checkbox"/> N/A <input type="checkbox"/> Notes: Section Conforms to Required Specifications: YES <input type="checkbox"/> NO <input type="checkbox"/>
Additional Testing:	<div style="text-align: right;">N/A <input type="checkbox"/></div> TC____Pass <input type="checkbox"/> Fail <input type="checkbox"/> TC____Pass <input type="checkbox"/> Fail <input type="checkbox"/> TC____Pass <input type="checkbox"/> Fail <input type="checkbox"/> TC____Pass <input type="checkbox"/> Fail <input type="checkbox"/> TC____Pass <input type="checkbox"/> Fail <input type="checkbox"/> Notes: Additional Testing Conforms to Expected Results: YES <input type="checkbox"/> NO <input type="checkbox"/>
System Pass or Fail:	
Performed By: (Sign/Date)	
Reviewed By: (Sign/Date)	

6.2 Review and Approval

This completed **Section 6**, identified as **Final Verification Report**, documents that Lab Kinetics Incubating 8-well Tube Reader and BG Analytics® software passed all specified testing in this System Verification Protocol and will adequately perform when used for the purposes for which it was intended.

Review and Approval	
_____	_____
Signature: Authorized Individual	Date

Title	
_____	_____
Signature: Authorized Individual	Date

Title	

7 Attachments

7.1 Training Records

7.2 Objective Evidence

7.3 Additional Testing

7.4 Discrepancy Report

7.5 Problem Resolution Report

7.6 Maintenance

Contact Information

Corporate Headquarters

Associates of Cape Cod, Inc.

124 Bernard E. Saint Jean Drive
East Falmouth, MA 02536-4445 USA
Tel: (888) 395-2221 or (508) 540-3444
Fax: (508) 540-8680
E-mail: custservice@acciusa.com
www.acciusa.com

United Kingdom

Associates of Cape Cod Int'l., Inc.


Deacon Park, Moorgate Road
Knowsley, Liverpool L33 7RX
United Kingdom
Tel: (44) 151-547-7444
Fax: (44) 151-547-7400
E-mail: info@acciuk.co.uk
www.acciuk.co.uk

Europe

Associates of Cape Cod Europe GmbH


Opelstrasse 14
D-64546 Mörfelden-Walldorf
Germany
Tel: (49) 61 05-96 10 0
Fax: (49) 61 05-96 10 15
E-mail: service@acciusa.de
www.acciusa.de

Authorized representative

 Emergo Europe, Prinsessegracht 20, 2514 AP, The Hague, The Netherlands

Note: serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.

Symbols Used

 Indicates compliance with the requirements of all the applicable EU directives

 In Vitro Diagnostic Device

 Product Model Name

 Manufacturer

 EU Representative

Revision History

Rev 2: Added Downloading procedure, Authorized representative, Revision History and Symbol used sections. Modified Section 5.3. Minor clarifications and formatting. Updated document name within quality system.

Rev 3: Modified sections 2.3, 3.4 and 3.5 for use of the new PKF08 Calibration Tool.